

Bid Document/ बिड दस्तावेज़

Bid Details/बिड विवरण	
Bid End Date/Time/बिड बंद होने की तारीख/समय	03-06-2024 10:00:00
Bid Opening Date/Time/बिड खुलने की तारीख/समय	03-06-2024 10:30:00
Bid Offer Validity (From End Date)/बिड पेशकश वैधता (बंद होने की तारीख से)	90 (Days)
Ministry/State Name/मंत्रालय/राज्य का नाम	Pmo
Department Name/विभाग का नाम	Department Of Atomic Energy
Organisation Name/संगठन का नाम	Directorate Of Purchase And Stores
Office Name/कार्यालय का नाम	Directorate Of Purchase And Stores
Total Quantity/कुल मात्रा	14
Item Category/मद केटेगरी	Digital Multimeter as per IS 9000 (Q3) , true rms clamp meter (Q3) , Insulation Tester (Q3) , Digital Tachometer (Q3) , LCR Meter (Q3)
MSE Exemption for Years of Experience and Turnover/ अनुभव के वर्षों से एमएसई छूट	No
Startup Exemption for Years of Experience and Turnover/ अनुभव के वर्षों से स्टार्टअप छूट	No
Document required from seller/विक्रेता से मांगे गए दस्तावेज़	Certificate (Requested in ATC) *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer
Bid to RA enabled/बिड से रिवर्स नीलामी सक्रिय किया	No
Type of Bid/बिड का प्रकार	Two Packet Bid
Primary product category	Digital Multimeter as per IS 9000
Time allowed for Technical Clarifications during technical evaluation/तकनीकी मूल्यांकन के दौरान तकनीकी स्पष्टीकरण हेतु अनुमत समय	2 Days
Inspection Required (By Empanelled Inspection Authority / Agencies pre-registered with GeM)	No
Evaluation Method/मूल्यांकन पद्धति	Item wise evaluation/

EMD Detail/ईएमडी विवरण

Required/आवश्यकता	No
-------------------	----

ePBG Detail/ईपीबीजी विवरण

Required/आवश्यकता	No
-------------------	----

Splitting/विभाजन

Bid splitting not applied/बोली विभाजन लागू नहीं किया गया.

MII Purchase Preference/एमआईआई खरीद वरीयता

MII Purchase Preference/एमआईआई खरीद वरीयता	No
--	----

MSE Purchase Preference/एमएसई खरीद वरीयता

MSE Purchase Preference/एमएसई खरीद वरीयता	Yes
---	-----

1. Purchase preference to Micro and Small Enterprises (MSEs): Purchase preference will be given to MSEs as defined in Public Procurement Policy for Micro and Small Enterprises (MSEs) Order, 2012 dated 23.03.2012 issued by Ministry of Micro, Small and Medium Enterprises and its subsequent Orders/Notifications issued by concerned Ministry. If the bidder wants to avail the Purchase preference, the bidder must be the manufacturer of the offered product in case of bid for supply of goods. Traders are excluded from the purview of Public Procurement Policy for Micro and Small Enterprises. In respect of bid for Services, the bidder must be the Service provider of the offered Service. Relevant documentary evidence in this regard shall be uploaded along with the bid in respect of the offered product or service. If L-1 is not an MSE and MSE Seller (s) has/have quoted price within L-1+ 15% (Selected by Buyer) of margin of purchase preference /price band defined in relevant policy, such Seller shall be given opportunity to match L-1 price and contract will be awarded for 25%(selected by Buyer) percentage of total QUANTITY. The buyers are advised to refer the OM No.F.1/4/2021-PPD dated 18.05.2023 [OM No.1 4 2021 PPD dated 18.05.2023](#) for compliance of Concurrent application of Public Procurement Policy for Micro and Small Enterprises Order, 2012 and Public Procurement (Preference to Make in India) Order, 2017.

2. Estimated Bid Value indicated above is being declared solely for the purpose of guidance on EMD amount and for determining the Eligibility Criteria related to Turn Over, Past Performance and Project / Past Experience etc. This has no relevance or bearing on the price to be quoted by the bidders and is also not going to have any impact on bid participation. Also this is not going to be used as a criteria in determining reasonableness of quoted prices which would be determined by the buyer based on its own assessment of reasonableness and based on competitive prices received in Bid / RA process.

Evaluation Method (Item Wise Evaluation Method)

Contract will be awarded schedulewise and the determination of L1 will be done separately for each schedule. The details of item-consignee combination covered under each schedule are as under:

Evaluation Schedules	Item/Category	Quantity
----------------------	---------------	----------

Schedule 1	Digital Multimeter As Per Is 9000	3
Schedule 2	True Rms Clamp Meter	3
Schedule 3	Insulation Tester	3
Schedule 4	Insulation Tester	1
Schedule 5	Digital Tachometer	1
Schedule 6	Lcr Meter	3

Digital Multimeter As Per IS 9000 (3 pieces)

Technical Specifications/तकनीकी विशिष्टियाँ

[* As per GeM Category Specification/जेम कैटेगरी विशिष्टि के अनुसार](#)

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
MEASUREMENTS	DC Voltage Range	1000.0 (Volt)
	DC Voltage Accuracy	0.1 (percent)
	DC Voltage resolution	0.1 (milliVolt)
	AC Voltage Range	1000.0 (Volt)
	AC Voltage Accuracy	0.5 (percent)
	AC Voltage resolution	0.1 (milliVolt)
	DC Current Range	10.0 (Ampère)
	DC Current Accuracy	1.0 (percent)
	DC Current resolution	0.001 (milliAmpère)
	AC Current Range	10.0 (Ampère)
	AC Current Accuracy	0.5 (percent)
	AC Current resolution	0.001 (milliAmpère)
	Resistance Range	60.0 (megaOhm)
	Resistance Accuracy	1.0 (percent)
	Resistance resolution	0.1 (Ohm)
	Capacitance Range (Micro farad)	.01 to 1000
	Capacitance Accuracy (%)	3.5
	Capacitance resolution (nF)	0.01
	Frequency Range (KHz)	100
	Frequency Accuracy (%)	0.1
	Frequency resolution (Hz)	0.01

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
	Temperature Range (Degree C)	-40 to 400
	Temperature Accuracy (%)	2
	Temperature resolution (Degree C)	0.1
FEATURES	Type	Hand Held
	Display (Digits)	3.5
	Measurement Type	True RMS
	Counts	6000, 50000, 10000, 20000, 30000, 40000, 60000, 70000, 80000, 90000, 100000, 200000, 20000000 Or higher
OTHER FEATURES	Continuity test and Diode test	Available
	Data logging	Available
	If Data Logging Feature is Available, Internal Memory	min. 5000 points/ 200 readings
	Computer Interface along with software	Available
SAFETY	Safety Rating	Category IV 600 V of IEC 61010
	Safety Certification	CE
POWER REQUIREMENTS	Power Requirements	Battery Operated
	if Battery Operated, then Type of Battery	Non- Rechargeable (Alkaline)
	if rechargeable , then chemistry of the battery	NA
	if Non-rechargeable , then size of the battery / cell	AA
	number of non- rechargeable the battery / cell	4
	Suitable charger / adaptor provided	NA
WARRANTY	Warranty	1.0 (year)
OPERATING CONDITION	Minimum Operating Temperature	-10.0 (degree Celsius)
	Maximum Operating Condition	55.0 (degree Celsius)
	Operating Humidity (RH)	80.0 (percent)
TEST REPORT DETAILS	The equipment supplied with calibration certificate	YES

Specification	Specification Name/विशिष्ट का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
	conformity to Dry Heat for 16 hrs at a specified maximum temp in accordance with IS:9000/part-3/section-5/1977 latest	Yes
	conformity to Cold Test for 4 hrs at a specified minimum temp in accordance with IS:9000/part-2/section-4/1977 latest	Yes
	conformity to Damp Heat (Cyclic)Test for 2 Cycles of 24 hrs each at a temp of 40 degree C & 95% RH	Yes
	Availability of Test Reports from Central Govt/NABL/ILAC approved accredited lab to prove conformity to the specification	Yes

Additional Specification Parameters - Digital Multimeter As Per IS 9000 (3 pieces)

Specification Parameter Name	Bid Requirement (Allowed Values)
TERMINAL SHUTTER	Terminal shutter has to be present to prevent incorrect test leads insertion in current terminals
FEATURES	Auto hold function, Minimum-Maximum-Average function, Continuity check, Diode test, Data logging, Auto power off, LCD backlit features have to be present
Accessories	Test Leads, Fuses (1 Set), calibration certificate, Carrying case have to be supplied alongwith the instrument.
Battery life	Minimum 200 hours of battery life should be present in DC voltage mode
Technical details	Bidder should state make, model and submit technical datasheet alongwith the bid

* Bidders offering must also comply with the additional specification parameters mentioned above.

Consignees/Reporting Officer/प्रेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्रा

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्रा	Delivery Days/डिलीवरी के दिन
1	Bendapudi Hemalatha	507116,STORES UNIT, HWP MANUGURU, P.O. GAUTAMINAGAR, ASWAPURAM, KHAMMAM	3	90

True Rms Clamp Meter (3 pieces)

Technical Specifications/तकनीकी विशिष्टियाँ

[* As per GeM Category Specification/जेम कैटेगरी विशिष्टि के अनुसार](#)

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
GENERAL FEATURES	Type	Hand Held
	Display	3.5 digit Backlit LCD
	Measurement Type	True RMS
	count	6000
MEASUREMENT FEATURES	Resistance Measurement Range	4.0 (kiloOhm)
	Resistance Measurement Accuracy (+/-)	1.0 (percent)
	Frequency Measurement Range	0.5 (kiloHertz)
	Frequency Measurement Accuracy (+/-)	0.5 (percent)
	AC Current Measurement Range	100.0 (Ampère)
	AC Current Measurement Accuracy (+/-)	2.0 (percent)
	AC Voltage Measurement Range	600.0 (Volt)
	AC Voltage Measurement Accuracy (+/-)	1.5 (percent)
	DC Current Measurement Range	100.0 (Ampère)
	DC Current Measurement Accuracy (+/-)	2.0 (percent)
	DC Voltage Measurement Range	400.0 (Volt)
	DC Voltage Measurement Accuracy (+/-)	1.0 (percent)
GENERIC	Jaw Size (Dia in Closed State)	30.0 (millimeter)

Specification	Specification Name/विशिष्ट क नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
POWER REQUIREMENTS	Power Requirements	Battery Operated
	Type of Battery	Non- Rechargeable (Alkaline)
	if rechargeable , then chemistry of the battery	NA
	if Non-rechargeable , then size of the battery / cell	AAA
	Suitable charger/adaptor provided	NA
OPERATING CONDITION	Minimum Operating Temperature	0.0 (degree Celsius)
	Maximum Operating Temperature	40.0 (degree Celsius)
	Operating Humidity (RH)	90.0 (percent)
SAFETY	Safety Certification	UL, CE
	Safety Rating	CAT III 600 V of IEC 61010 -2-D31
SERVICE	Warranty	2.0 (year)
TEST REPORT DETAILS & TESTS	The equipment supplied with calibration certificate	yes
	conformity to Dry Heat for 16 hrs at a specified maximum temp	No
	conformity to Cold Test for 4 hrs at a specified minimum temp in accordance with IS:9000/part-2/section-4/1977 latest	No
	conformity to Damp Heat (Cyclic)Test for 2 Cycles of 24 h each at a temp of 40 degree C & 95% RH	No
	Availability of Test Reports from Central Govt/NABL/ILAC approved accredited lab to prove conformity to the specification	Yes

Additional Specification Parameters - True Rms Clamp Meter (3 pieces)

Specification Parameter Name	Bid Requirement (Allowed Values)
Features	Zero function feature to be present which allows the display to be cleared for DC measurements, Data hold function has to be present
Technical Details	Bidder should state make, model and submit technical datasheet along with the bid

* Bidders offering must also comply with the additional specification parameters mentioned above.

Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्रा

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्रा	Delivery Days/डिलीवरी के दिन
1	Bendapudi Hemalatha	507116, STORES UNIT, HWP MANUGURU, P.O. GAUTAMINAGAR, ASWAPURAM, KHAMMAM	3	90

Insulation Tester (3 pieces)

Technical Specifications/तकनीकी विशिष्टियाँ

* As per GeM Category Specification/जेम कैटेगरी विशिष्टि के अनुसार

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
PERFORMANCE	Type	Hand Held
	Measurement Range	0.01 M Ohm to 10 G Ohm
	Measurement Accuracy	2.0 (percent)
	Measurement Resolution	0.01 (megaOhm)
	Test Voltage	1000.0 (Volt)
	Test Current	1.0 (milliAmpère)
OTHER MEASUREMENT	AC/DC Voltage Measurement (Volt)	600
	AC/DC Voltage Measurement Resolution (Volts)	0.1
	AC/DC Voltage Measurement Accuracy (%)	2.0
FEATURES	Pass Fail Function	Yes
	Data Logging	NA
	PC interface along with suitable & driven software	Not Available

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
	Live Circuit Detection	inhibit test terminal voltage >30 V prior to initialization of test
SAFETY	Safety rating	Category IV 600 V of IEC 61010
POWER REQUIREMENT	Power Requirements	Battery Operated
	Type of Battery	Non- Rechargeable (Alkaline)
	if rechargeable , then chemistry of the battery	NA
	if Non-rechargeable , then size of the battery / cell	AA
	Number of non- rechargeable the battery / cell	4
WARRANTY	Warranty	1.0 (year)
OPERATING CONDITION	Minimum Operating Temperature	-20.0 (degree Celsius)
	Maximum Operating Temperature	55.0 (degree Celsius)
	Operating Humidity (RH)	95.0 (percent)
TEST REPORT DETAILS	The equipment supplied with calibration certificate	YES
	conformity to Dry Heat for 16 hrs at a specified maximum temp	No
	conformity to Cold Test for 4 hrs at a specified minimum temp in accordance with IS:9000/part-2/section-4/1977 latest	No
	conformity to Damp Heat (Cyclic)Test for 2 Cycles of 24 hours each at a temp of 40 degree C & 95% RH	No
	Availability of Test Reports from Central Govt/NABL/ILAC approved accredited lab to prove conformity to the specification	Yes

Additional Specification Parameters - Insulation Tester (3 pieces)

Specification Parameter Name	Bid Requirement (Allowed Values)
Features	Features of Auto power off, backlit display, auto-discharge of capacitive voltage for added user protection have to be present. Should automatically calculate Polarisation index and Dielectric Absorption ratio.
Battery Life	Battery life of atleast 750 insulation tests
Technical details	Bidder should state make, model and submit technical datasheet alongwith the bid. Test leads, calibration certificate, carrying case have to be supplied alongwith the instrument
Test voltages	250V, 500V and 1000V

* Bidders offering must also comply with the additional specification parameters mentioned above.

Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्रा

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्रा	Delivery Days/डिलीवरी के दिन
1	Bendapudi Hemalatha	507116,STORES UNIT, HWP MANUGURU, P.O. GAUTAMINAGAR, ASWAPURAM, KHAMMAM	3	90

Insulation Tester (1 pieces)

Technical Specifications/तकनीकी विशिष्टियाँ

[* As per GeM Category Specification/जेम केटेगरी विशिष्टि के अनुसार](#)

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
PERFORMANCE	Type	Hand Held
	Measurement Range	0.01 M Ohm to 10 G Ohm
	Measurement Accuracy	2.0 (percent)
	Measurement Resolution	0.1 (megaOhm)
	Test Voltage	5000.0 (Volt)
	Test Current	1.0 (milliAmpère)
OTHER MEASUREMENT	AC/DC Voltage Measurement (Volt)	NA
	AC/DC Voltage Measurement Resolution (Volts)	NA
	AC/DC Voltage Measurement Accuracy (%)	NA
FEATURES	Pass Fail Function	No

Specification	Specification Name/विशिष्ट का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
	Data Logging	NA
	PC interface along with suitable & driven software	Not Available
	Live Circuit Detection	NA
SAFETY	Safety rating	Category III 600 V of IEC 61010
POWER REQUIREMENT	Power Requirements	Hand Driven
	Type of Battery	NA
	if rechargeable , then chemistry of the battery	NA
	if Non-rechargeable , then size of the battery / cell	NA
	Number of non- rechargeable the battery / cell	NA
WARRANTY	Warranty	1.0 (year)
OPERATING CONDITION	Minimum Operating Temperature	0.0 (degree Celsius)
	Maximum Operating Temperature	50.0 (degree Celsius)
	Operating Humidity (RH)	90.0 (percent)
TEST REPORT DETAILS	The equipment supplied with calibration certificate	YES
	conformity to Dry Heat for 16 hrs at a specified maximum temp	No
	conformity to Cold Test for 4 hrs at a specified minimum temp in accordance with IS:9000/part-2/section-4/1977 latest	No
	conformity to Damp Heat (Cyclic)Test for 2 Cycles of 24 hours each at a temp of 40 degree C & 95% RH	No
	Availability of Test Reports from Central Govt/NABL/ILAC approved accredited lab to prove conformity to the specification	No

Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्रा

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्रा	Delivery Days/डिलीवरी के दिन
1	Bendapudi Hemalatha	507116,STORES UNIT, HWP MANUGURU, P.O. GAUTAMINAGAR, ASWAPURAM, KHAMMAM	1	90

Digital Tachometer (1 pieces)**Technical Specifications/तकनीकी विशिष्टियाँ**

[* As per GeM Category Specification/जेम केटेगरी विशिष्टि के अनुसार](#)

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
Battery	Battery type	2-AA batteries
Additional Information	Accessories	Reflectors, self-adhesive
Test Reports	Availability of test report from central Govt/ NABL/ILAC accredited lab covering all parameters	Yes
	Copies of test Reports/certificates and other certifications to be furnished to buyer on demand at time of supplies	Yes

Additional Specification Parameters - Digital Tachometer (1 pieces)

Specification Parameter Name	Bid Requirement (Allowed Values)
Range	Range: 0 to 3000 rpm and Accuracy: +- 0.02% reading +1
Sensing distance	Sensing distance upto 500mm
Battery life	Battery life of minimum 30 hours
Features	Backlit display, Maximum/minimum/average value, auto power off features have to be present

Specification Parameter Name	Bid Requirement (Allowed Values)
Technical details	Calibration certificate, Min. ten no. of reflective strips have to be supplied alongwith the instrument.

* Bidders offering must also comply with the additional specification parameters mentioned above.

Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्रा

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्रा	Delivery Days/डिलीवरी के दिन
1	Bendapudi Hemalatha	507116,STORES UNIT, HWP MANUGURU, P.O. GAUTAMINAGAR, ASWAPURAM, KHAMMAM	1	90

LCR Meter (3 pieces)

Technical Specifications/तकनीकी विशिष्टियाँ

[* As per GeM Category Specification/जेम केटेगरी विशिष्टि के अनुसार](#)

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
Certification & standard	Manufacturer Certification	ISO 9001(2015) or Latest
	Compliance	CE
Product Information	Equivalent Circuit	Parallel and Series
	Input impedance AC and DC Voltage	1M ohm
	Measurement parameters	Capacitance (C), Inductance (L), Resistance (R)
	Storage	Vinyl Pouch
	Features	Backlight Display, Comparator functions
	Basic Accuracy L (rdg-Reading, dgt-Digit)	+/- (0.7% rdg+5 dgt)
	Test fixture/Test Leads	Built-in test fixture
	Correction Function(Calibration Function)	Open
	Ranging Mode	Manual
	Measurement Range Duty Cycle	NA
	Maximum counts of LCR Display	1999

Specification	Specification Name/विशिष्ट का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
	Measurement Range Frequency	2KHz to 15MHz
	Basic Accuracy R (rdg-Reading, dgt-Digit)	+/- (0.5%rdg+3dgt)
	Maximum counts of Q.D.R Display	1999
	No. of Digits in LCR Display	3.5
	Measurement Range DC Voltage	0-10V
	Display	Digital LCD Display
	Measurement Range AC Current	20mA to 10A
	Measurement Range Q	0.001 to 1999
	Measurement Range D	0.001 to 1999
	Measurement Range L	2000microH to 10000H
	Basic Accuracy C (rdg-Reading, dgt-Digit)	+/- (0.7% rdg+3 dgt)
	Measurement Range C	2000pf to 10mF
	Test Frequency (in Hertz)	8.2
	External Power Source	DC power adapter
	Measurement Range AC Voltage	200mV to 750V
	Internal Power Source	NA
	Test Parameters	LS+(Q,D,RS)
	No. of Digits in Q.D.R Display	NA
	ACCESSORIES	Test leads pair
	Measurement Range DC Current	NA
	Basic Resolution L	0.001 (microHenry)
	Basic Resolution C	0.1 (picoFarad)
	Basic Resolution R	1m ohm
	Display digit Size mm	14mm
	PC connectivity interface	RS-232 interface
	Measurement Range ? (Phase Angle)	0.00 degree to +/- 180.0 degree
	Measurement Range Z, R, X	20 ohm to 10M ohm
Warranty	Warranty (min)	1, 2, 3 Or higher (year)

Specification	Specification Name/विशिष्ट का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
Test Report	Test Report To Be Submitted To The Buyer On Demand	Yes
	Availability of Test Report from Central Govt/NABL/ILAC accredited lab to prove conformity to specification	Yes

Additional Specification Parameters - LCR Meter (3 pieces)

Specification Parameter Name	Bid Requirement (Allowed Values)
Features	Capacitive discharge protection, Auto power off, capacitor with the function of self discharge features have to be present
Accessories	Carrying case, test leads have to be supplied alongwith the instrument.

* Bidders offering must also comply with the additional specification parameters mentioned above.

Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्रा

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्रा	Delivery Days/डिलीवरी के दिन
1	Bendapudi Hemalatha	507116,STORES UNIT, HWP MANUGURU, P.O. GAUTAMINAGAR, ASWAPURAM, KHAMMAM	3	90

Buyer Added Bid Specific Terms and Conditions/क्रेता द्वारा जोड़ी गई बिड की विशेष शर्तें

1. Scope of Supply

Scope of supply (Bid price to include all cost components) : Only supply of Goods

2. Warranty

Warranty period of the supplied products shall be as given in specifications from the date of final acceptance of goods or after completion of installation, commissioning & testing of goods (if included in the scope of supply), at consignee location. OEM Warranty certificates must be submitted by Successful Bidder at the time of delivery of Goods. The seller should guarantee the rectification of goods in case of any break down during the guarantee period. Seller should have well established Installation, Commissioning, Training, Troubleshooting and Maintenance Service group in INDIA for attending the after

sales service. Details of Service Centres near consignee destinations are to be uploaded along with the bid.

3. **Certificates**

Bidder's offer is liable to be rejected if they don't upload any of the certificates / documents sought in the Bid document, ATC and Corrigendum if any.

4. **Buyer Added Bid Specific ATC**

Buyer Added text based ATC clauses

1. REF. NO. DPS/RPUM/G1/ 22920

2. Bidders are required to provide a self-declaration certificate stating their status as manufacturer/trader for the tendered item, along with their bid, failing which their bid will not be considered for MSE purchase preference.

3. Any bidder from a country which shares a land border with India shall be eligible to bid in this tender only if the bidder is registered with the Department for promotion of Industry and Internal Trade (DPIIT). The bidders shall enclose valid registration certificate along with their offer. Wherever the bids are received without accompanying the above said requisite certificate such offers shall be treated as incomplete and not considered. Furthermore, every bidder participating against this Department tender shall invariably enclose along with the Bid, **a self-declared undertaking "Annexure to Bid Form: Eligibility Declarations (Annexure-XII)", duly mentioning the tender reference of this office as stated above, failing which Bid will be summarily rejected.**

5. **Buyer Added Bid Specific ATC**

Buyer uploaded ATC document [Click here to view the file.](#)

Disclaimer/अस्वीकरण

The additional terms and conditions have been incorporated by the Buyer after approval of the Competent Authority in Buyer Organization, whereby Buyer organization is solely responsible for the impact of these clauses on the bidding process, its outcome, and consequences thereof including any eccentricity / restriction arising in the bidding process due to these ATCs and due to modification of technical specifications and / or terms and conditions governing the bid. If any clause(s) is / are incorporated by the Buyer regarding following, the bid and resultant contracts shall be treated as null and void and such bids may be cancelled by GeM at any stage of bidding process without any notice:-

1. Definition of Class I and Class II suppliers in the bid not in line with the extant Order / Office Memorandum issued by DPIIT in this regard.
2. Seeking EMD submission from bidder(s), including via Additional Terms & Conditions, in contravention to exemption provided to such sellers under GeM GTC.
3. Publishing Custom / BOQ bids for items for which regular GeM categories are available without any Category item bunched with it.
4. Creating BoQ bid for single item.
5. Mentioning specific Brand or Make or Model or Manufacturer or Dealer name.
6. Mandating submission of documents in physical form as a pre-requisite to qualify bidders.
7. Floating / creation of work contracts as Custom Bids in Services.
8. Seeking sample with bid or approval of samples during bid evaluation process. (However, in bids for [attached categories](#), trials are allowed as per approved procurement policy of the buyer nodal Ministries)
9. Mandating foreign / international certifications even in case of existence of Indian Standards without specifying equivalent Indian Certification / standards.

10. Seeking experience from specific organization / department / institute only or from foreign / export experience.
11. Creating bid for items from irrelevant categories.
12. Incorporating any clause against the MSME policy and Preference to Make in India Policy.
13. Reference of conditions published on any external site or reference to external documents/clauses.
14. Asking for any Tender fee / Bid Participation fee / Auction fee in case of Bids / Forward Auction, as the case may be.

Further, if any seller has any objection/grievance against these additional clauses or otherwise on any aspect of this bid, they can raise their representation against the same by using the Representation window provided in the bid details field in Seller dashboard after logging in as a seller within 4 days of bid publication on GeM. Buyer is duty bound to reply to all such representations and would not be allowed to open bids if he fails to reply to such representations.

[This Bid is also governed by the General Terms and Conditions/ यह बिड सामान्य शर्तों के अंतर्गत भी शासित है](#)

In terms of GeM GTC clause 26 regarding Restrictions on procurement from a bidder of a country which shares a land border with India, any bidder from a country which shares a land border with India will be eligible to bid in this tender only if the bidder is registered with the Competent Authority. While participating in bid, Bidder has to undertake compliance of this and any false declaration and non-compliance of this would be a ground for immediate termination of the contract and further legal action

in accordance with the laws./जेम की सामान्य शर्तों के खंड 26 के संदर्भ में भारत के साथ भूमि सीमा साझा करने वाले देश के बिडर से खरीद पर प्रतिबंध के संबंध में भारत के साथ भूमि सीमा साझा करने वाले देश का कोई भी बिडर इस निविदा में बिड देने के लिए तभी पात्र होगा जब वह बिड देने वाला सक्षम प्राधिकारी के पास पंजीकृत हो। बिड में भाग लेते समय बिडर को इसका अनुपालन करना होगा और कोई भी गलत घोषणा किए जाने व इसका अनुपालन न करने पर अनुबंध को तत्काल समाप्त करने और कानून के अनुसार आगे की कानूनी कार्रवाई का आधार होगा।

---Thank You/धन्यवाद---